

Product Information Sheet

申請書編號(如適用)Application No.(If applicable): eye _____
 家居電話號碼(如適用)Residential Telephone No.(If applicable): _____
 收據編號(如適用)Sales Memo No.(If applicable): _____

| | |
|--|---|
| 血糖機 / Blood Glucose Meter | |
| 牌子 / Brand | 拜耳 / Bayer |
| 型號 / Model | 拜安康 / Contour TS |
| 原產地 / Place of Manufacture | 日本 / Japan |
| 主要功能 / Key features | <ul style="list-style-type: none"> ❖ 免調碼技術可消除錯誤調碼所帶來的問題 ❖ 只需 0.6 μl 極微量血液樣本, 8 秒測試時間 ❖ 250 組結果記憶 ❖ No Coding TM technology - eliminate accuracy and usability issues due to miscoding ❖ 0.6 microliter small sample size, 8 second test time ❖ 250 test memory |
| 配套 / Packaging List | 1 x 說明書 / 1 x Instruction manual 1 x 採血器 / 1 x Lancing device 1 x 血糖機袋 / 1 x Blood glucose meter pouch |
| 產品價格/ Price | 協意價格請參照銷售備忘錄/發票 Agreed as per Sales Memo/Invoice |
| 附加費 / Additional Delivery Charge | 指定地區/ 地址 (如離島) 有可能需付額外送貨附加費 Additional delivery charge may be added for specific area/ location (e.g. outlying island) |
| 保修人身份/ Identity of repair service provider | 尚健醫療/ Celki Medical |
| 保修地址/ Repair service address | Tel: 8100-6386 , email: diabetescare@bayerhealthcare.com |
| 換貨政策 / Exchange policy | 只適用於銷售備忘錄/發票日期 10 天內 (由 HKT 提供)。相關條款及條件詳情請參閱銷售備忘錄/發票背頁。 Only within 10 days of the date of Sales Memo (provided by HKT). Please refer to the terms and conditions on the back of the Sales Memo for details. |
| 保修期 / Warranty period | 主機 Main: 永久保養/ lifetime warranty 配件 Accessory: 沒有保養/ No warranty |

上述產品資料摘錄自生產商/供應商文獻及/或由生產商/供應商所提供之資料

The above information is extracted from literature and / or information on this product by manufacturer / supplier

Hong Kong Telecommunications (HKT) Limited
 查詢熱線 Enquiry hotline: 2888 0008
<http://www.pccw-hkt.com>

Product Information Sheet 20130726